

Test Report

To Marumuraseicha Co.,Ltd.

SEIKAN KENSA CENTER Inc.
2310 Takayanagi, Fujieda, Shizuoka, 426-0041, JAPAN
TEL.:+81-54-634-1000 FAX:+81-54-634-1010



The result of the examination is reported as mentioned in the following.

1. Sample Information

Sample name	2015 production Green tea
Sampling date	***
Sampled by	***
Date Sample Received	February 22, 2016

2. Measuring method

Nuclide analysis method by gamma spectrometry with germanium semiconductor detector

"Radioactive caesium in the food inspection method(2012)" Ministry of Health, Labour and Welfare(Japan)

3. Result

Item	Result	Detection limit
Radioactive Iodine(I-131)	N.D.	2.5Bq/kg
Radioactive Caesium(Cs-134)	N.D.	2.6Bq/kg
Radioactive Caesium(Cs-137)	N.D.	2.7Bq/kg

Note: Germanium Detector Model : CANBERRA GC4020_DSA-1000(MCA)

Measured Date : February,22 2016

"N.D.(Not Detected)"in the result means that the measured value is under the method of detection limit.

***** End of Test Report *****